



# UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE  
United States Patent and Trademark Office  
Address: COMMISSIONER FOR PATENTS  
P.O. Box 1450  
Alexandria, Virginia 22313-1450  
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/716,437	11/20/2003	Tomaso Vercellotti	2247-114	6624
6449	7590	10/25/2005	EXAMINER	
ROTHWELL, FIGG, ERNST & MANBECK, P.C. 1425 K STREET, N.W. SUITE 800 WASHINGTON, DC 20005			PATEL, NIHIR B	
			ART UNIT	PAPER NUMBER
			3743	

DATE MAILED: 10/25/2005

Please find below and/or attached an Office communication concerning this application or proceeding.

SP

<b>Office Action Summary</b>	<b>Application No.</b> 10/716,437	<b>Applicant(s)</b> VERCELLOTTI ET AL.	
	<b>Examiner</b> Nihir Patel	<b>Art Unit</b> 3743	

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --  
**Period for Reply**

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If the period for reply specified above is less than thirty (30) days, a reply within the statutory minimum of thirty (30) days will be considered timely.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

#### Status

- 1) ☒ Responsive to communication(s) filed on September 28<sup>th</sup>, 2005.
- 2a) ☐ This action is FINAL. 2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

#### Disposition of Claims

- 4) ☐ Claim(s) \_\_\_\_\_ is/are pending in the application.
- 4a) Of the above claim(s) \_\_\_\_\_ is/are withdrawn from consideration.
- 5) ☐ Claim(s) \_\_\_\_\_ is/are allowed.
- 6) ☒ Claim(s) 9, 12, 15 and 16 is/are rejected.
- 7) ☐ Claim(s) \_\_\_\_\_ is/are objected to.
- 8) ☐ Claim(s) \_\_\_\_\_ are subject to restriction and/or election requirement.

#### Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on \_\_\_\_\_ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.  
 Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).  
 Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

#### Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some \* c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
  2. ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_\_.
  3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).
- \* See the attached detailed Office action for a list of the certified copies not received.

#### Attachment(s)

- |  |  |
|--|--|
| 1) <input checked="" type="checkbox"/> Notice of References Cited (PTO-892)  | 4) <input checked="" type="checkbox"/> Interview Summary (PTO-413)<br>Paper No(s)/Mail Date. _____ |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948)                                   | 5) <input type="checkbox"/> Notice of Informal Patent Application (PTO-152)                        |
| 3) <input type="checkbox"/> Information Disclosure Statement(s) (PTO-1449 or PTO/SB/08)<br>Paper No(s)/Mail Date _____ | 6) <input type="checkbox"/> Other: _____   |

Art Unit: 3743

**DETAILED ACTION**

***Response to Arguments***

Applicant's arguments, see interview summary with respect to the advisory action filed on September 21<sup>st</sup>, 2005 have been fully considered and are persuasive. The advisory action has been withdrawn.

***Claim Rejections - 35 USC § 103***

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

Claims 9, 12, 15 and 16 are rejected under 35 U.S.C. 103(a) as being unpatentable over Mehta et al. (US 5,843,109). Referring to claims 9, 12, 15 and 16, the applicant's specification does not establish any criticality on why the instrument is used for extraction of impacted third molars in the vicinity of the dental alveolus; for vertebral laminectomy treatment; for hand and foot bone surgery and therefore considers it a matter of design choice. It would have been obvious to one of ordinary skill in the art at the time of the invention to provide an instrument that can be used for different situations, based on the teachings of Mehta (column 1 lines 59-67) "Hand-piece apparatus in accordance with the present invention useful for the disruption (ie, fragmentation, eroding, sloughing off and emulsification) and removal of unwanted material such as tissue, cartilage, bone calculi or the like from an animal body such as a human...". The usage of the instrument is based on patient criteria (ie, hand, foot, skull, etc...) and therefore is matter of design choice on where to use the instrument.


Art Unit: 3743

***Conclusion***

The prior art made of record and not relied upon is considered pertinent to applicant's disclosure.

Any inquiry concerning this communication or earlier communication from the examiner should be directed to Nihir Patel whose telephone number is (571) 272-4803. The examiner can normally be reached on Monday-Friday from 7:30 am to 4:30 pm. If attempts to reach the examiner by telephone are unsuccessful the examiner supervisor Henry Bennett can be reached at (571) 272 4791.

Nihir Patel  
October 18<sup>th</sup>, 2005

  
Henry Bennett  
Supervisor, Patent Examiner  
Group 3700